

Calibration in practice

Vincent Piétu
(IRAM)

Outline

I. The Plateau de Bure interferometer

II. On-line calibrations

III. CLIC

IV. Off-line calibrations

Foreword



An **automated data reduction pipeline** exists for Plateau de Bure data. It is run automatically at Bure on all observed projects, and helps the **astronomer on duty** to assess data quality, project completeness, etc.

It also contains many useful informations for the **visitor** coming to IRAM to reduce data, and hence a first step is often to look at its results prior to really “reduce” data.

This talk will be illustrated with plots that can be found in the pipeline (especially the “**show**”, that displays system parameters), so you get familiar with it. This will be indicated with ***Pipeline***.

I. The Plateau de Bure interferometer:

Antennas and stations

Receivers

Signal transport


IF Processor

Correlators

The Plateau de Bure interferometer: Antennas

- 6 antennas (on alt-az mounts) ...
- ... that can be put on 32 stations ...
- ... on 3 arms (W, N, E).

The interferometer arms



The Plateau de Bure interferometer: Antennas

6 antennas (on alt-az mounts) ...
... that can be put on 32 stations ...
... on 3 arms (W, N, E).

Each ...
... composed of 216 panels.
... weighting 130 tons
... measuring 15 m in diameter

Bure antenna with snow



Antennas in compact configuration



The Plateau de Bure interferometer: Receivers

Equipped with 3 (soon 4) receiver bands ...


... which are **Single-Side Band** (SSB), with sideband rejection of the order of 10-20 dB.


... and **dual polarization** (orthogonal linear polarizations, but quarter-wave plate available in Band 1).

... observing **2 x 4 GHz**

... converting this bandwidth to an Intermediate Frequency (IF) of frequency **4-8 GHz**.

Receiver:





The Plateau de Bure Interferometer: Transporting the signal

The electromagnetic incident wave has been converted in a **electric analogical** signal in the receiver (with a **phase relation** between the two).

The down-conversion is done by mixing the astronomical signal with a **local oscillator** (LO) which is a monochromatic wave with controlled phase.







The LOs in the different antennas are all generated from a common **frequency synthesizer** (located in the central building), and this frequency reference is transported through High-Q coaxial cables.

The **pathlength of which** has to be monitored


Pipeline: monitoring the cable phase.


RF: Uncal. CLIC - 30-SEP-2010 09:39:48 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
Ph: Abs. (39 3311 P CORR)-(593 3769 P CORR) 20-JUN-2010 04:57-10:59

Scan Avg.
WIDEX Unit 1



The IF-processor





“Narrow-band” correlator (a.k.a CAMEMBERT)




High-spectral resolution to zoom with enhanced spectral capabilities.

And the wide-band correlator (a.k.a. WIDEX)




Large bandwidth to cover all the receiver bandwidth and increase sensitivity

Quantization of the signal





		$S(x)$			
		0	0	1	1
		$M(x)$			
		1	0	0	1
$S(y)$	$M(y)$				
0	1	n^2	n	$-n$	$-n^2$
0	0	n	1	-1	$-n$
1	0	$-n$	-1	1	n
1	1	$-n^2$	$-n$	n	n^2

Quantization of the signal



Why adjusting sampling levels ? (tweaking ...)









Pipeline: monitoring the tweak levels

Narrow: gain in the [1-256] range

RF: Uncal. CLIC - 30-SEP-2010 09:27:18 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
 Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Abs. (39 3311 P CORR)-(593 3769 P CORR) 20-JUN-2010 04:57-10:59

Scan Avg.
WIDEX Unit 1









Pipeline: monitoring the tweak levels

WIDEX: attenuation in the [1-64] range

RF: Uncal. CLIC - 30-SEP-2010 09:28:56 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
 Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Abs. (39 3311 P CORR)-(593 3769 P CORR) 20-JUN-2010 04:57-10:59

Scan Avg.
WIDEX Unit 1



II.a. On-line calibrations:

Atmospheric calibration

Pointing

Focusing

Measuring the instrumental delay

Atmospheric calibration.

This is essential to convert the output of the correlators (counts) to a temperature scale.

This requires:







- Determining the system noise (**receiver temperature**).
- Determining the single-dish **gain**.
- Determining the **atmospheric absorption**.
- Linearity of the receiving system.

Chopper wheel method + use of an **atmospheric model** (**ATM**, *Cernicharo & Pardo*)

Measuring the receiver temperature

RF: Uncal. CLIC - 29-SEP-2010 14:22:01 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320)V Q3(320,320,320)H
Ph: Abs. (579 3766 P CALI)-(581 3766 P CALI) 20-JUN-2010 10:56-10:57







No Avg.
WIDEX Unit 1



Pipeline: monitoring the receiver temperature (vs time)

RF: Uncal. CLIC - 30-SEP-2010 09:18:19 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
 Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Abs. (39 3311 P CORR)-(593 3769 P CORR) 20-JUN-2010 04:57-10:59







Scan Avg.
WIDEX Unit 1



Pipeline: monitoring the receiver temperature (vs freq.)

RF: Uncal. CLAS - 29-SEP-2010 15:09:33 - pietu@ahop-pietu W08W05E03W11N07N02 6Dq
 Am: Abs. WIDE CH3GGH 158.230GHz B2 Q3(320,320,320,320)M D3(320,320,320,320)H
 Ph: Abs. (580 3766 P CAL) -(580 3766 P CAL) 20-JUN-2010 10:57-10:57

No Avg.
 PHYSICAL ant.
 WDMX Unit 4



Atmospheric calibration: outputs







As a result of atmospheric calibration, we derive:

- The **receiver temperature**.
- The amount of **water vapor** (expressed as “precipitable water vapor”).
- The **system temperature** (*i.e.* the total noise of the atmosphere+telescope).

Pipeline: monitoring calibration parameters

RF: Uncal. CLIC - 30-SEP-2010 09:20:44 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
 Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Abs. (39 3311 P CORR)-(593 3769 P CORR) 20-JUN-2010 04:57-10:59







Scan Avg.
WIDEX Unit 1




Pipeline: monitoring calibration parameters

RF: Uncal. CLIC - 30-SEP-2010 09:21:27 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
 Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Abs. (39 3311 P CORR)-(593 3769 P CORR) 20-JUN-2010 04:57-10:59


Scan Avg.
WIDEX Unit 1



Pointing




Pointing



Pipeline: monitoring the pointing corrections

RF: Uncal. CLIC - 29-SEP-2010 13:41:36 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
 Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Abs. (39 3311 P CORR)-(584 3769 P CORR) 20-JUN-2010 04:57-10:59


Scan Avg.
VERTICAL pol.



Focusing ...

Subband used: C10 C12


Scan: 3306 R2 Source: 30454.3 S= 1.00 Az= -78.16 El= 34.63



Focusing ...

Subband used: C10 C12







Scan: 3306 R2 Source: 30454.3 S= 1.00 Az= -78.16 El= 34.63



Measuring the instrumental delay ...

RF: Uncal. CLIC - 29-SEP-2010 13:51:41 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
Ph: Abs. (1 3307 P GAIN)-(1 3307 P GAIN) 20-JUN-2010 04:48-04:48

Scan Avg.
VERTICAL pol.



Measuring the instrumental delay ...

```
I-SOLVE_DELAY,[3307] Delay offset for Phys.Ant. 1 : 0.000 +- 0.000 ns
I-SOLVE_DELAY,[3307] Absolute delay for Log.Ant. 1 : 0.000 +- 0.000 ns
I-SOLVE_DELAY,[3307] 2 Ant. 2 Ch. L01 L02 L03 L04 L10 L12 Band LSB rms 7.304
I-SOLVE_DELAY,[3307] delay= 0.042 ns. phase= -5.533

I-SOLVE_DELAY,[3307] Delay offset for Phys.Ant. 2 : -0.042 +- 0.001 ns
I-SOLVE_DELAY,[3307] Absolute delay for Log.Ant. 2 : 242.958 +- 0.001 ns
I-SOLVE_DELAY,[3307] 3 Ant. 3 Ch. L01 L02 L03 L04 L10 L12 Band LSB rms 5.501
I-SOLVE_DELAY,[3307] delay= -0.050 ns. phase= 166.848

I-SOLVE_DELAY,[3307] Delay offset for Phys.Ant. 3 : 0.050 +- 0.000 ns
I-SOLVE_DELAY,[3307] Absolute delay for Log.Ant. 3 : 556.741 +- 0.000 ns
I-SOLVE_DELAY,[3307] 4 Ant. 4 Ch. L01 L02 L03 L04 L10 L12 Band LSB rms 5.049
I-SOLVE_DELAY,[3307] delay= 0.156 ns. phase= -0.798
```

[...]

Most of the delays come from the fiber optics.

III. The CLIC Software (Continuum and Line Interferometric Calibration)

Introduction

Data format

Useful commands

CLIC: part of GILDAS

Part of **GILDAS** (Grenoble Image and Line Data Analysis Software), developed and maintained mainly in Grenoble.

GILDAS composed of:

- **Kernel:**
 - SIC: command interpreter, computer
 - GREG: graphics
- **Packages:**
 - e.g. CLASS
 - CLIC

CLIC: what is CLIC ?

CLIC is able to **read/write file** with a data format **specific** to the PdBI.

CLIC is able to **plot various quantities** stored (or derived from) in the data file.

CLIC is able to do various type of **fits to the data**, necessary for data/system calibration.

CLIC is able to **store these corrections**.

CLIC is able to **export uv tables** in the GILDAS .uvt format.

Data format

At the PdBI, an **IPB file (.ipb or .IPB)** is written:

- Collection of observations related to a **single project** (“track”).
- Contains
 - An **index** of the observations.
 - **Observations** themselves.
- **Binary format** (metadata and data).

Those file are transferred to the **database** in Grenoble and archived (can be retrieved through the **getproj** command).

Data format

Observations contain:

- **Observation header** composed of section containing observations parameters (frequency plan, calibration parameters, source information etc ...)
- Data composed of **records** each having
 - A **data header** (for parameter changing each second).
 - The **data** (working with **logical** numbering).
- The records are:
 - **Temporal**: one dump every second (spectral averaged). Referred to as the continuum subbands: **C01 [...] C12**.
 - **Spectral**: two spectra at the end of the integration period (time averaged). Referred to as the line subbands: **L01 [...] L12**.

CLIC: miscellaneous

All information stored either in the observation header or the data headers can be accessed through **SIC variables** than can be used (and changed !) afterward

This allows the use of **procedures** which are of higher levels than the basic commands (called when clicking on widgets).

Calibration is done on a **.hpb file**, which contains **only the observations header** (among which the **calibration sections**).

CLIC, as part of GILDAS, has automatic **keyword completion**. System commands are called with **\$**, procedures with **@**. Command options start with **/**.

CLIC: basic commands

Although use of widgets (hence procedure) is recommended, it may be required to know some commands.

HELP: help on CLIC or a specific command.

FIND: allow to build an index of observations (on which we will apply commands, e.g. plotting, fitting, storing ...).

LIST: list the content of the index.

SET: sets parameters (e.g. SET X TIME ; SET Y PHASE or SET SUBBANDS L01 to L04 ; SET AVER SCAN).

PLOT: plot the observations in the index.

SOLVE: make a fit to data (e.g. SOLVE PHASE /PLOT).

TABLE: writes a uv table.

III. Offline calibrations: procedures & widgets

Select

Autoflag

Phcorr

RF

Phase

Flux

Amplitude

Select

Open the **hpb file**.

Find if there is source observations.

Find the used **receiver band** and sky **sideband**.

Find if configuration changed during observations.

Determine if receiver re-tuning (new GAIN scan).

Find the better **bandpass (RF) calibrator**.

Find the amp/phase calibrators.

Create internal (SIC) variables used by the subsequent procedures.

Useful variables

Variables created and updated after select:

'do_atm': enable/disable radiometric phase correction

'band_source': calibrator used for RF calibration

'phcal': calibrator for amp/pha calibration

'do_avpol': average pol (or not) for amp calibration

Can be overridden with **let** (e.g. `let do_atm .false.`)

Autoflag

Check for **hardware/software failure** (by comparing observing date with a database with known problems).

Check for possible **timing error** (scan too long/too short or UT update problem).

Check if source observations surrounded by **flagged calibrator** observations, and flag data if needed.

Determine if the phase-corrected (by use of the Water Vapor Radiometer at 22 GHz) data are better than the uncorrected data.


By **comparing amplitude** on the calibrator of the corrected and uncorrected data. Correction applied to the sources if closest (in time) calibrator found to be better.

Can be bypassed with:

STORE CORRECTION BAD|GOOD /ANT n

Check for possible interference in one of the 3 bands of the WVR (possible interference by Hotbird 6).

Avoiding Hotbird 6



Phcorr - ctd

Check if amplitude calibrator polarized.

Phcorr - ctd


RF: Uncal. CLIC - 04-OCT-2010 12:36:06 - pietu@pctcp02 N11W05W08N07 5Dq-E03 T.A.: 60. s.
 Am: Abs. RB56 C054R 84.527GHz B1 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Abs. Wvr. (111 4508 P CORR)-(1167 5370 P CORR) 05-SEP-2007 04:47-15:51 0954+658

DOP.1: 10.1%

DOP.2: 10.3%

DOP.3: 10.4%

DOP.4: 10.4%



RF: Radio Frequency bandpass calibration

Goal: calibrate the radio-frequency bandpass (Intermediate frequency already calibrated by mean of observations of a noise source – IFPB in file).


Assumption: no temporal dependance.

How:

- observations of a strong quasar
- Self-calibration and averaging
- fit of polynoms as a function of frequency (leaving the average amp/pha unchanged) by **antenna** or baseline
- End precision needed depends on projects


RF calibration

RF: Uncal. CLIC - 04-OCT-2010 08:49:18 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
 Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Rel.(A) Atm. (38 3310 P FLUX)-(43 3315 P CORR) 20-JUN-2010 04:56-05:01



RF calibration

RF: Uncal. CLIC - 04-OCT-2010 08:52:28 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320)V Q3(320,320,320)H
Ph: Rel.(A) Atm. (38 3310 P FLUX)-(43 3315 P CORR) 20-JUN-2010 04:56-05:01



Phase calibration

Goal: correct for **temporal variations** of

- Electronics
- Local oscillators
- Antenna position or time errors
















... and **estimate atmospheric phase noise**

How: by using observations of unresolved calibrators

- Plot the quasar phase
 - Should be zero if coordinates are precise enough
- Fit a **spline** to the **antenna or the baseline gains**
 - SET PHASE ANTENNA|BASELINE
 - Possibility to use polynoms (SOLVE PHA /POL degree)
- Store correction (scan based).

Phase calibration

RF: Fr.(A) CLIC - 04-OCT-2010 08:59:43 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
Am: Abs. WIDE CH3CCH 159.230GHz B2 Q3(320,320,320)V Q3(320,320,320)H
Ph: Abs. Atm. (96 3342 P FLUX)-(593 3769 P CORR) 20-JUN-2010 05:30-10:59



Flux calibration

Critical point of the calibration !

Needs the *a priori* knowledge of at least one observed calibrator's flux.







How:

- **Fix the flux** of one calibrator (Jy).
- Derive the **antenna efficiencies** by dividing the fixed flux (Jy) by the observed antenna temperature (K).
- Select the 3 best antenna (lowest Jy/K)
- Use efficiencies to **derive calibrators fluxes** (K x Jy/K).
- Store calibrator fluxes

Flux calibration

RF: Fr.(A) CLIC - 04-OCT-2010 09:18:59 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
 Am: Scaled WIDE CH3COH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
 Ph: Rel.(A) Atm. (38 3310 P FLUX)-(593 3769 P CORR) 20-JUN-2010 04:56-10:59

Scan Avg.
BOTH polarizations



Amplitude calibration

Goal: correct for **temporal variations** of

- Atmospheric **decorrelation**
- Antenna pointing/focusing
- **Antenna efficiency** (deicing on/off, etc ...)

By using observations of unresolved calibrators

- Plot the quasar amplitude (T_a^*) divided by their flux (Jy).

This is the inverse of the antenna efficiency

- Fit a **spline** to the **antenna or the baseline gains**.

















SET PHASE ANTENNA|BASELINE

Possibility to use polynoms (SOLVE AMP /POL degree)

- Store correction (scan based).

Amplitude calibration

RF: Fr.(A) CLIC - 04-OCT-2010 09:31:55 - pietu@dhcp-pietu W08W05E03N11N07N02 6Dq
Am: Scaled WIDE CH3CCH 159.230GHz B2 Q3(320,320,320,320)V Q3(320,320,320,320)H
Ph: Rel.(A) Atm. (96 3342 P FLUX)-(593 3769 P CORR) 20-JUN-2010 05:30-10:59



Creating a *uv* table

GILDAS *uv tables* (derived from GILDAS images).

Done with command **TABLE**

- Interface between CLIC and MAPPING
- Internal binary format
- Header+visibilities

Can be converted to fits if needed for use in other softwares.

Otherwise, next step occurs in MAPPING with imaging and deconvolution, or for uv-plane analysis.

Creating a *uv* table

Two modes:

- **Continuum**: produces one visibility per scan/baseline/correlator input
- **Line**: produces a spectra per scan/baseline

User selects data to be used

- **SET SELECTION CONT|LINE LSB|USB|DSB L01 TO L04**

Apply calibrations according to user choice

- **SET ANTENNA RELATIVE|ABSOLUTE ANT|BASE ATM|NOATM**
- **SET AMPLITUDE [...] JANSKY|KELVIN**
- **SET RF ON|OFF FILE|MEMORY**

Conclusions

All that is fine, but nothing's worth a good **tutorial** !